

re: WT Docket No. 01-309

As a hearing aid manufacturer, we strenuously object to any efforts to change FCC's decision regarding wireless phone compatibility with hearing aids.

The ANSI C63.19 test is an approved ANSI standard that we believe is repeatable, predictable and accurate.

We are already working to ensure that our hearing aids achieve at least a U2 rating under the ANSI C63.19 test that predicts compatibility with the U3-rated handsets that telephone manufacturers are required by the new rules to produce.